

RESEARCH

Open Access



Efficiency of a mechanical device in controlling tracheal cuff pressure in intubated critically ill patients: a randomized controlled study

Saad Nseir^{1,2*}, Andrey Rodriguez¹, Paula Saludes¹, Julien De Jonckheere³, Jordi Valles¹, Antonio Artigas¹ and Ignacio Martin-Loeches⁴

Abstract

Background: Cuff pressure (P_{cuff}) control is mandatory to avoid leakage of oral secretions passing the tracheal tube and tracheal ischemia. The aim of the present trial was to determine the efficacy of a mechanical device (PressureEasy®) in the continuous control of P_{cuff} in patients intubated with polyvinyl chloride (PVC)-cuffed tracheal tubes, compared with routine care using a manometer.

Methods: This is a prospective, randomized, controlled, cross-over study. All patients requiring intubation with a predicted duration of mechanical ventilation ≥ 48 h were eligible. Eighteen patients randomly received continuous control of P_{cuff} with PressureEasy® device for 24 h, followed by discontinuous control (every 4 h) with a manual manometer for 24 h, or vice versa. P_{cuff} and airway pressure were continuously recorded. P_{cuff} target was 25 cmH₂O during the two periods.

Results: The percentage of time spent with P_{cuff} 20–30 cmH₂O (median (IQR) 34 % (17–57) versus 50 % (35–64), $p = 0.184$) and the percentage of time spent with $P_{\text{cuff}} < 20$ cmH₂O (23 % (5–63) versus 43 % (16–60), $p = 0.5$) were similar during continuous control of P_{cuff} and routine care, respectively. However, the percentage of time spent with $P_{\text{cuff}} > 30$ cmH₂O was significantly higher during continuous control compared with routine care of tracheal cuff (26 % (14–39) versus 7 % (1–18), $p = 0.002$). No significant difference was found in P_{cuff} (25 (18–28) versus 21 (18–26), $p = 0.17$), mean airway pressure (14 (10–17) versus 14 (11–16), $p = 0.679$), or coefficient of variation of P_{cuff} (19 % (11–26) versus 20 % (11–25), $p = 0.679$) during continuous control compared with routine care of tracheal cuff, respectively.

Conclusions: PressureEasy® did not demonstrate a better control of P_{cuff} between 20 and 30 cmH₂O, compared with routine care using a manometer. Moreover, the device use resulted in significantly higher time spent with overinflation of tracheal cuff, which might increase the risk for tracheal ischemic lesions.

Trial registration: Clinicaltrial.gov: NCT02109003

Keywords: Cuff pressure; Control; Tracheal ischemia; Microaspiration; Mechanical ventilation

* Correspondence: s-nseir@chru-lille.fr

¹Critical Care Center, Hospital de Sabadell, Universidad Autonoma de Barcelona, CIBER enfermedades respiratorias, Corporació Sanitaria Parc Tauli, Parc Tauli 1, 08208 Sabadell, Spain

²Critical Care Center, R. Salengro Hospital, CHRU de Lille, rue E. Laine, 59037 Lille Cedex, France

Full list of author information is available at the end of the article

Background

Intubation is an invasive procedure that is still performed in a large proportion of critically ill patients [1]. Some long-term intubation-related complications are caused by inappropriate cuff pressure (P_{cuff}), and include microaspiration, and tracheal ischemic lesions [2, 3]. Microaspiration of contaminated oropharyngeal secretions is the main route of entry for bacteria into the lower respiratory tract [4, 5]. Colonization of the lower respiratory tract could progress into ventilator-associated pneumonia (VAP) when local or general defense mechanisms are altered in intubated critically ill patients [6]. VAP is the most frequent infection acquired in the ICU and is associated with high morbidity and mortality, especially in patients with comorbidities [7]. Tracheal ischemic injury is also associated with high morbidity, especially when routine care for tracheal cuff is not adequately provided [8–11].

Based on international recommendations, P_{cuff} should be kept between 20 and 30 cmH₂O using a manometer [12, 13]. However, several recent studies suggested that a noncontinuous control of P_{cuff} using a manometer is not effective in intubated critically ill patients [14]. Further, previous studies did not identify modifiable risk factors for overinflation or underinflation of tracheal cuff [15–18].

Recently, several devices allowing continuous control of P_{cuff} were evaluated [19–22]. These devices are classified into pneumatic (i.e., does not require power supply, but a single-use 200-ml cylindrical cuff) and electronic devices (requiring power supply). Two prospective randomized controlled animal and human studies first validated the use of a pneumatic device [19, 20]. A non-commercially available device has also been validated and proved to be efficient in continuously controlling P_{cuff} in ICU patients [22]. However, other commercially available automated devices were only validated in *in vitro* studies [23, 24]. Recent data suggest that these devices interfere with the self-sealing characteristics of high-volume low-pressure (HVLP) tracheal cuffs [23]. Further, these electronic devices have been shown to be less efficient than the pneumatic device in continuous control of P_{cuff} because of rapid correction of overinflation episodes [21].

A new mechanical device with an improved design aiming at continuously controlling P_{cuff} has been commercialized (PressureEasy®, Smiths medical). The advantages of using such device are its small size and lower cost compared with other devices. However, to our knowledge, no study has evaluated the efficiency of this device in controlling P_{cuff} . Therefore, we conducted this prospective randomized cross-over study to determine the efficiency of this device in continuously controlling P_{cuff} . Our hypothesis was that this mechanical device

would allow significant reduction of time spent with underinflation or overinflation of P_{cuff} , compared with routine care, using a manual manometer.

Methods

Ethical aspects

This prospective randomized controlled study was performed in a 16-bed ICU of the teaching hospital of Sabadell (Spain), from April 2014 to July 2014. The institutional review board of the Parc Tauli University Hospital approved the study. Written consent was obtained from the patients or from their next of kin (ClinicalTrials.gov identifier: NCT02109003 <http://clinicaltrials.gov/show/NCT02109003>).

Inclusion and exclusion criteria

All patients older than 18 years who were intubated with a predicted duration of mechanical ventilation ≥ 48 h were eligible for the study. Patients were excluded if they (1) were admitted to the ICU with a previous tracheostomy, (2) were enrolled in another study that might influence this study results, or (3) were intubated and mechanically ventilated >48 h at the time of randomization.

Randomization

Patients were randomly assigned to receive continuous control of P_{cuff} with the mechanical device (PressureEasy®) (Fig. 1) for 24 h, followed by discontinuous control (every 4 h) with a manual manometer (Hi-Lo Hand Pressure Gauge, Covidien, TM, Malinckrodt TM) for 24 h; or discontinuous control of P_{cuff} followed by continuous control of P_{cuff} . The target of P_{cuff} was 25 cmH₂O during the two periods. Randomization was performed using a computer-generated random assignment list in balanced blocs of four. Treatment assignments were contained in sealed opaque envelopes sequentially numbered.

PressureEasy® is a single-patient use device, designed to continuously monitor tracheal P_{cuff} . Its indicator window signals that P_{cuff} is maintained between 20 and 30 cmH₂O. In addition, the airway pressure auto-feedback feature boosts P_{cuff} to ensure proper sealing when high pressures are used during ventilation. A pressure feedback line is designed to eliminate cuff leaks at peak inspiratory pressure.

Outcome measurement

In order to determine the percentage of patients with underinflation and overinflation of tracheal cuff and the duration of these episodes, the P_{cuff} and the airway pressure were continuously recorded at a digitizing frequency of 100 Hz for 48 h (Physiotrace®, Estaris, Lille, France) [25], including 24 h of continuous control of




Fig. 1 The PressureEasy® device. **a** Tracheal tube, **(b)** ventilator circuit, **(c)** PressureEasy®, and **(d)** external tracheal cuff

P_{cuff} using the mechanical device and 24 h of manual control of P_{cuff} using the manometer. The connection between the pressure transducer and the tracheal cuff was identical in the two study periods, with a three-way stopcock of which the third port was either connected to the mechanical device or closed and connected to the manometer every 4 h (Fig. 2). A second pressure transducer was connected to the heat and moisture exchanger, in order to record respiratory pressure. Pressure transducers were connected to a laptop, in order to continuously transfer pressure-time curves and mean values. Connections were checked every 4 h. The engineer who analyzed the data (JDJ), using the same program, was blinded to the randomization order.

Study population

All the patients were intubated with a high-volume low-pressure PVC-cuffed tracheal tube with continuous subglottic secretion drainage (TaperGuard Evac, Covidien, Mallinckrodt) and were included within the first 48 h of intubation. Tracheal tube size was 8 and 7.5 in men and women, respectively. During the manometer period, nurses adjusted P_{cuff} every 4 h. Tracheal suctioning was performed six times a day or more frequently if clinically indicated.

Other definitions

Underinflation of tracheal cuff was defined as $P_{\text{cuff}} < 20$ cmH₂O for >5 min over the 24-h period of recording. Overinflation of tracheal cuff was defined as $P_{\text{cuff}} > 30$ cmH₂O for >5 min over the 24-h period of recording [14].

The primary objective was to determine the efficiency of the mechanical device in reducing the percentage of time spent with underinflation or overinflation of tracheal cuff, compared with routine care using a manometer. The secondary objective was to determine the efficiency of the mechanical device in reducing the percentage of patients with underinflation or overinflation of tracheal cuff, compared with routine care.

The coefficient of variation of P_{cuff} was calculated as standard deviation/mean $P_{\text{cuff}} \times 100$.

Statistical analyses

Sample size calculation

Based on unpublished preliminary results, the mean percentage of time spent with underinflation or overinflation of tracheal cuff was 30 % (SD 20 %) in patients intubated with a PVC-cuffed tracheal tube receiving routine care of P_{cuff} using a manual manometer. The expected mean percentage of time with underinflation or overinflation of tracheal cuff using the mechanical device was 10 % (expected difference of 20 %). Considering a power of 90 % and an alpha risk of 5 %, the inclusion of 22 patients was required in a parallel-group design. However, when we took into account the cross-over design, the number of patients to include was 18.

Result analysis

All analyses were performed in an intention-to-treat manner. Distribution of quantitative variables was tested using the Shapiro-Wilk test. Normally and nonnormally distributed variables were expressed as mean \pm SD and median (25th, 75th interquartile), respectively. The




Fig. 2 Description of connections between pressure transducers, ventilator, and tracheal tube. **a** Tracheal tube, **(b)** connection between external cuff and the three-way stopcock, **(c)** ventilator, **(d)** three-way stopcock, **(e)** PressureEasy® device or manometer, **(f)** pressure transducers connected to the ventilator and to tracheal cuff, and **(g)** laptop

difference between the two groups was considered as significant if $p < 0.05$. McNemar and Wilcoxon tests were used to compare qualitative and quantitative variables between the two 24-h periods, respectively.

We calculated the median (25th, 75th interquartile) value of mean airway pressure in all patients and considered patients with airway pressure >75th interquartile as patients with high airway pressure. In order to determine the impact of high airway pressure on the efficiency of the mechanical device, we repeated all statistical analyses after exclusion of patients with high airway pressure.

Results

Patient characteristics

Eighteen patients were included in this study. Their mean age was 62 ± 12 years. At ICU admission, SAPS II and SOFA scores were 51 ± 14 and 7.7 ± 3.1 , respectively. Duration of mechanical ventilation and ICU stay were 25 ± 18 days and 33 ± 21 days, respectively. No significant difference was found between the two study periods

regarding the percentage of patients receiving sedation or neuromuscular-blocking agents. Glasgow Coma score, SOFA score, and PEEP were also similar during the two periods (Table 1). All patients received assist-control ventilation during the recording period. No significant difference was found in mean airway pressure between the two study periods (Table 2).

Impact of the mechanical device on P_{cuff}

No significant difference was found in mean P_{cuff} , percentage of time spent with P_{cuff} between 20 and 30 cmH₂O, percentage of time spent with $P_{\text{cuff}} < 20$ cmH₂O, or coefficient of variation of P_{cuff} . The percentage of time spent with $P_{\text{cuff}} > 30$ cmH₂O was significantly higher during continuous control compared with routine care of tracheal cuff (Table 2). No significant difference was found in the percentage of patients with P_{cuff} 20–30 cmH₂O, $P_{\text{cuff}} < 20$ cmH₂O, or $P_{\text{cuff}} > 30$ cmH₂O during continuous control of P_{cuff} and routine care (Table 2). No significant difference was found in the number of

Table 1 Patient characteristics during the 48 h following randomization

Variables	Continuous control of P_{cuff} $n = 18$	Routine care $n = 18$	p values
SOFA score	8.8 ± 3.4	9 ± 3.7	0.458
Sedation	14 (77)	14 (77)	>0.999
Neuromuscular-blocking agent use	1 (5)	0 (0)	>0.999
Glasgow coma score ^a	7.5 ± 3	9 ± 3.7	0.137
PEEP	6.7 ± 1.9	6.6 ± 1.9	0.157

Data are frequencies (%) or mean \pm SD

SOFA sequential organ failure assessment, PEEP positive end-expiratory pressure

^aThe verbal response was evaluated as in intubated patients: 1 no answer, 2 seems able to give simple responses, and 5 seems able to speak

underinflation episodes between continuous control and routine care periods (median (IQR) 0 (0, 5) versus 2 (0, 7), $p = 0.733$).

Subgroup analysis

After exclusion of the four patients with high mean airway pressure (>17 cmH₂O), no significant difference was found in mean P_{cuff} (26 (18, 28) versus 22 (19, 27) cmH₂O, $p = 0.177$), percentage of time spent with P_{cuff} between 20 and 30 cmH₂O (50 (24, 58) versus 53 (24, 71), $p = 0.124$), percentage of time spent with $P_{\text{cuff}} < 20$ cmH₂O (16 (2, 51) versus 35 (9, 51), $p = 0.3$), or

Table 2 Impact of the mechanical device on tracheal cuff pressure

Variables	Continuous control of P_{cuff} $n = 18$	Routine care $n = 18$	p values
Mean airway pressure, cmH ₂ O	14 (10–17)	14 (11–16)	0.679
Mean P_{cuff} , cmH ₂ O	25 (18, 28)	21 (18, 26)	0.17
Coefficient of variation of P_{cuff} , %	19 (11, 26)	20 (12, 25)	0.67
P_{cuff} 20–30 cmH ₂ O			
Yes ^a	18 (100)	18 (100)	NA
% of time ^b	34 (17, 57)	50 (35, 64)	0.184
$P_{\text{cuff}} < 20$, cmH ₂ O			
Yes ^a	17 (94)	18 (100)	>0.999
% of time ^b	23 (5, 63)	43 (16, 60)	0.528
$P_{\text{cuff}} > 30$, cmH ₂ O			
Yes ^a	18 (100)	16 (89)	0.500
% of time ^b	26 (14, 40)	7 (1, 18)	0.002

Data are frequencies (%) or median (interquartile range)

 P_{cuff} cuff pressure, NA not applicable^aYes indicates the number of patients with at least one P_{cuff} 20–30, >20, or >30 cmH₂O^b% of time indicates all the time spent with P_{cuff} 20–30, >20, or >30 cmH₂O reported to the total recording time during each study period

coefficient of variation of P_{cuff} (15 (10, 22) versus 16 (10, 24), $p = 0.363$). The percentage of time spent with $P_{\text{cuff}} > 30$ cmH₂O was still significantly higher during continuous control compared with routine care of tracheal cuff (31 (18, 42) versus 8 (3, 20), $p = 0.006$).

Discussion

The results of our study suggest that PressureEasy® was not more efficient in maintaining P_{cuff} within the recommended range (20–30 cmH₂O) than routine care of tracheal cuff using a manometer every 4 h. No significant difference was found between continuous control of P_{cuff} using the mechanical device and routine care regarding P_{cuff} , percentage of time spent with underinflation of tracheal cuff, and coefficient of variation of P_{cuff} . However, the percentage of time spent with overinflation of tracheal cuff was significantly higher during continuous control compared with routine care.

One potential explanation for this result is the high frequency of P_{cuff} control using the manometer, i.e., every 4 h, during routine care. A recent prospective study performed in a cohort of 102 patients receiving invasive mechanical ventilation and manual control of P_{cuff} using a manometer every 8 h found a lower percentage of time (18 %) spent within the target range [14]. However, this explanation is unlikely because our results are in agreement with those of another recent trial, in which P_{cuff} was controlled using a manometer every 8 h [13]. Further, previous studies showed that each time a manometer is connected to the tracheal cuff, a sudden drop of P_{cuff} is observed [22, 26]. This is probably related to transient underinflation of tracheal cuff and might promote microaspiration of contaminated secretions [27]. Therefore, controlling P_{cuff} more frequently in order to maintain it within the target range could probably not be recommended.

In spite of strictly applying instructions for user of the mechanical device, the percentage of time spent within the target range of P_{cuff} was only 34 %. Previous studies using different devices to continuously control P_{cuff} reported better performances and a percentage within the targeted range >90 % [19, 20, 22, 28]. Another potential explanation for the low efficiency of the device is the mechanism of P_{cuff} control. This device uses the respiratory flow to inflate tracheal cuff during inspiration. Therefore, it is dependent on respiratory flow and airway pressure that may widely vary in critically ill patients. In addition, the key principle is probably an equilibrium between airway pressure and P_{cuff} , which means that the time constant of the system might play a role in the higher P_{cuff} [29, 30]. Further, the exact P_{cuff} target, using the mechanical device, could not be precisely determined or modified. In order to test the hypothesis that our results could be explained by the bad performance

of the device in patients with high airway pressures, we repeated all statistical analyses after exclusion of patients with the highest mean airway pressure but found similar results. This might be explained by the fact that high airway pressure could have occurred during only a short period of the total recording time and the relatively small number of included patients.

To our knowledge, our study is the first to evaluate the PressureEasy® mechanical device, and its results suggest that the device should not be used in critically ill patients. Our study also raises the important question of why medical devices such as tracheal tubes or P_{cuff} controllers could obtain the *Communauté Européenne* (CE) mark and be used in critically ill patients without any published clinical data proving their efficiency. Nevertheless, clinical evaluation before CE mark could be very complex and might reduce the development of new technologies.

No significant difference was found in coefficient of variation of P_{cuff} between the two study periods. However, P_{cuff} variation in study patients was relatively small and had probably no clinical impact, except in those patients with P_{cuff} around 20 or 30 cmH₂O.

Some limitations of our study should be acknowledged. First, this was a single-center study. All study patients received assist-control ventilation, and a high proportion of them were sedated. Therefore, our results may not be generalized to patients in other ICUs, especially to nonsedated patients receiving pressure support ventilation. Second, the number of studied patients was relatively small. However, P_{cuff} was continuously recorded for 48 h. In addition, the number of included patients was calculated based on our hypothesis before starting the study. Third, because of the study design, namely the fact that every patient was its own control, we could not evaluate complications related to underinflation or overinflation of tracheal cuff. However, we think that this design is probably optimal as a first step in order to validate the device, before performing larger studies evaluating its impact on complications. Our study design allowed adjustment of patient-related confounding factors, such as tracheal anatomy, tracheal tube size, and airway pressure. Fourth, we excluded patients who could not be included during the first 48 h of their invasive mechanical ventilation. This exclusion criterion was selected because duration of intubation was identified as a risk factor for underinflation of tracheal cuff [14]. Fifth, we did not collect P_{cuff} values obtained manually during the routine care period and did not evaluate the relationship between manometer connection and any drop in P_{cuff} . However, several previous well-designed and performed studies have clearly confirmed this relationship [22, 26]. Finally, Glasgow Coma score was used to evaluate

consciousness in study patients. However, a sedation score would have been more appropriate in sedated patients.

Conclusions

The PressureEasy® device did not demonstrate a better control of P_{cuff} within the target range (20–30 cmH₂O), compared with routine care using a manometer every 4 h. Moreover, the percentage of time spent with overinflation of tracheal cuff was more frequent using this device compared with routine care, which might increase the risk for tracheal ischemic lesions. Therefore, the use of this device could not be recommended in critically ill patients. Further large studies are required to confirm our results.

Abbreviations

CE: Communauté Européenne; HPLV: high-volume low-pressure; P_{cuff} : cuff pressure; PEP: positive expiratory pressure; SAPS: simplified acute physiology score; SOFA: sequential organ failure assessment; PVC: polyvinyl chloride.

Competing interests

SN is a member of the advisory board in Covidien. The other authors declare that they have no competing interests.

Authors' contributions

SN, JV, AA, and IML designed the study. AR and PS included the patients, performed the pressure recording, and collected the data. JDJ performed the data analyses. SN and IML coordinated the study. SN performed the statistical analysis and drafted the manuscript. All authors participated in the data interpretation and read and approved the final manuscript.

Author details

¹Critical Care Center, Hospital de Sabadell, Universidad Autonoma de Barcelona, CIBER enfermedades respiratorias, Corporació Sanitaria Parc Tauli, Parc Tauli 1, 08208 Sabadell, Spain. ²Critical Care Center, R. Salengro Hospital, CHRU de Lille, rue E. Laine, 59037 Lille Cedex, France. ³Clinical Investigation Center—Innovative Technologies, INSERM CIC-IT 807, University Hospital of Lille, 152 rue du Dr. Yersin, 59120 Loos, France. ⁴Multidisciplinary Intensive Care, St James's University Hospital, Dublin, Ireland.

Received: 6 January 2015 Accepted: 20 May 2015

Published online: 02 June 2015

References

- Esteban A, Frutos-Vivar F, Muriel A, Ferguson ND, Peñuelas O, Abaira V, et al. Evolution of mortality over time in patients receiving mechanical ventilation. *Am J Respir Crit Care Med*. 2013;188:220–30.
- Touat L, Fournier C, Ramon P, Salleron J, Durocher A, Nseir S. Intubation-related tracheal ischemic lesions: incidence, risk factors, and outcome. *Intensive Care Med*. 2013;39:575–82.
- Nseir S, Zerimech F, Jaillette E, Artru F, Balduyck M. Microaspiration in intubated critically ill patients: diagnosis and prevention. *Infect Disord Drug Targets*. 2011;11:413–23.
- Blot SI, Poelaert J, Kollef M. How to avoid microaspiration? A key element for the prevention of ventilator-associated pneumonia in intubated ICU patients. *BMC Infect Dis*. 2014;14:119.
- Beuret P, Philippon B, Fabre X, Kaaki M. Effect of tracheal suctioning on aspiration past the tracheal tube cuff in mechanically ventilated patients. *Ann Intensive Care*. 2012;2:45.
- Craven DE, Chroneou A, Zias N, Hjalmarson KI. Ventilator-associated tracheobronchitis: the impact of targeted antibiotic therapy on patient outcomes. *Chest*. 2009;135:521–8.
- Nair GB, Niederman MS. Ventilator-associated pneumonia: present understanding and ongoing debates. *Intensive Care Med*. 2014;31:34–48.
- Hameed AA, Mohamed H, Al-Mansoori M. Acquired tracheoesophageal fistula due to high intracuff pressure. *Ann Thorac Med*. 2008;3:23–5.

9. Seegobin RD, van Hasselt GL. Endotracheal cuff pressure and tracheal mucosal blood flow: endoscopic study of effects of four large volume cuffs. *Br Med J*. 1984;288:965–8.
10. Combes X, Schaulvliege F, Peyrouset O, Motamed C, Kirov K, Dhonneur G, et al. Intracuff pressure and tracheal morbidity: influence of filling with saline during nitrous oxide anesthesia. *Anesthesiology*. 2001;95:1120–4.
11. Jaillette E, Martin-Loeches I, Artigas A, Nseir S. Optimal care and design of the tracheal cuff in the critically ill patient. *Ann Intensive Care*. 2014;4:7.
12. Guidelines for the management of adults with hospital-acquired, ventilator-associated, and healthcare-associated pneumonia. *Am J Respir Crit Care Med* 2005, 171:388–416
13. Nseir S, Zerimech F, De Jonckheere J, Alves I, Balduyck M, Durocher A. Impact of polyurethane on variations in tracheal cuff pressure in critically ill patients: a prospective observational study. *Intensive Care Med*. 2010;36:1156–63.
14. Nseir S, Brisson H, Marquette C-H, Chaud P, Di Pompeo C, Diarra M, et al. Variations in endotracheal cuff pressure in intubated critically ill patients: prevalence and risk factors. *Eur J Anaesthesiol*. 2009;26:229–34.
15. Lizy C, Swinnen W, Labeau S, Poelaert J, Vogelaers D, Vandewoude K, et al. Cuff pressure of endotracheal tubes after changes in body position in critically ill patients treated with mechanical ventilation. *Am J Crit Care*. 2014;23:e1–8.
16. Sole ML, Su X, Talbert S, Penoyer DA, Kalita S, Jimenez E, et al. Evaluation of an intervention to maintain endotracheal tube cuff pressure within therapeutic range. *Am J Crit Care*. 2011;20:109–17. quiz 118.
17. Rubes D, Klein AA, Lips M, Rulisek J, Kopecky P, Blaha J, et al. The effect of adjusting tracheal tube cuff pressure during deep hypothermic circulatory arrest: a randomised trial. *Eur J Anaesthesiol*. 2014;31:452–6.
18. Servin SO, Barreto G, Martins LC, Moreira MM, Meirelles L, Neto JAC, et al. Atraumatic endotracheal tube for mechanical ventilation. *Rev Bras Anesthesiol*. 2011;61:311–9.
19. Duguet A, D'Amico L, Biondi G, Prodanovic H, Gonzalez-Bermejo J, Similowski T. Control of tracheal cuff pressure: a pilot study using a pneumatic device. *Intensive Care Med*. 2007;33:128–32.
20. Nseir S, Duguet A, Copin M-C, De Jonckheere J, Zhang M, Similowski T, et al. Continuous control of endotracheal cuff pressure and tracheal wall damage: a randomized controlled animal study. *Crit Care*. 2007;11:R109.
21. Brisson H, Bouhemad B, Lu Q, ROUBY JJ. Comparison of two automated endotracheal tube cuff pressure regulators in intubated critically ill patients. (abstract). *Intensive Care Med*. 2011;37:296.
22. Farré R, Rotger M, Ferré M, Torres A, Navajas D. Automatic regulation of the cuff pressure in endotracheally-intubated patients. *Eur Respir J*. 2002;20:1010–3.
23. Weiss M, Doell C, Koepfer N, Madjdpour C, Woitzek K, Bernet V. Rapid pressure compensation by automated cuff pressure controllers worsens sealing in tracheal tubes. *Br J Anaesth*. 2009;102:273–8.
24. Chenelle CT, Oto J, Sulemanji D, Fisher DF, Kacmarek RM. Evaluation of an automated endotracheal tube cuff controller during simulated mechanical ventilation. *Respir Care*. 2015;60:183–90.
25. De Jonckheere J, Logier R, Dassonneville A, Delmar G, Vasseur C. PhysioTrace: an efficient toolkit for biomedical signal processing. *Conf Proc Annu Int Conf IEEE Eng Med Biol Soc IEEE Eng Med Biol Soc Annu Conf*. 2005;7:6739–41.
26. Memela ME, Gopalan PD. Variations in endotracheal tube cuff pressure: is 8-hourly monitoring enough? *South African J Crit Care*. 2014;30:35–40.
27. Nseir S, Zerimech F, Fournier C, Lubret R, Ramon P, Durocher A, et al. Continuous control of tracheal cuff pressure and microaspiration of gastric contents in critically ill patients. *Am J Respir Crit Care Med*. 2011;184:1041–7.
28. Jaillette E, Zerimech F, De Jonckheere J, Makris D, Balduyck M, Durocher A, et al. Efficiency of a pneumatic device in controlling cuff pressure of polyurethane-cuffed tracheal tubes: a randomized controlled study. *BMC Anesthesiol*. 2013;13:50.
29. Bolzan DW, Gomes WJ, Peixoto TCA, Faresin SM, Carvalho AC, De Paola AAV, et al. Clinical use of the volume-time curve for endotracheal tube cuff management. *Respir Care*. 2014;59:1628–35.
30. Richard J-CM, Mercat A. Tracheal cuff management as part of a lung-protective strategy. *Respir Care*. 2014;59:1810–1.

Submit your manuscript to a SpringerOpen[®] journal and benefit from:

- Convenient online submission
- Rigorous peer review
- Immediate publication on acceptance
- Open access: articles freely available online
- High visibility within the field
- Retaining the copyright to your article

Submit your next manuscript at ► springeropen.com